

E0175 Application of T2 Candida test for diagnosis of candidaemia in real life

Emilio Cendejas³, Alba Ruiz*¹, Iker Falces Romero³, Anabel Aleixandre^{2 5}, Marcelo Fernandez¹, Julio Garcia Rodriguez³, Javier Peman Garcia⁴

¹Medical Research Institute La Fe, Severe Infection Research Group, Valencia, Spain, ²Hospital La Fe, Department of Surgical Critical Care, Valencia, Spain, ³Hospital Universitario La Paz, Department of Clinical Microbiology, Madrid, Spain, ⁴Hospital La Fe, Department of Clinical Microbiology, Valencia, Spain, ⁵La Fe University Hospital, Anesthesiology and Reanimation, Valencia, Spain

Background: An early and appropriated antifungal therapy is clue to improve the clinical outcome of patients with candidemia. T2Candida® Panel (T2 Biosystems) is rapid diagnostic approach (3-5 hours) that enables sensitive and specific detection of 5 *Candida* species (*C. albicans*/*C. tropicalis*, *C. glabrata*/*C. krusei* and *C. parapsilosis*) directly in whole blood. We compared the T2Candida results performed in two institutions with blood culture (BC) results in order to correlate them with the current gold standard in the clinical setting.

Materials/methods: Samples were obtained from patients with risks factors for candidemia admitted in La Fe University Hospital (Valencia, Spain) and La Paz University Hospital (Madrid, Spain). Informed consent from patients was obtained. Samples evaluated by the T2 Candida panel and BC using BacT/ALERT PLUS® VIRTUO™ were drawn simultaneously. T2Candida results were compared to BC.

Results: Sixty-five samples belonging to 48 patients were evaluated. Forty-eight blood samples were negative by the two methods. Ten and seven T2Candida tests were positive (Table) and invalid, respectively. Three out of five discordant results belonged to the same patient in whom the candidemia episode follow up was performed by T2Candida.

Patient	T2 <i>Candida</i> spp. detected	Blood culture result	Other <i>Candida</i> isolated (samples)
1	<i>C. albicans/tropicalis</i>	Negative	<i>C. albicans</i> (Peritoneal fluid)
2 (1/day 0)	<i>C. parapsilosis</i>	Negative	
2 (2/day +8)	<i>C. parapsilosis</i>	<i>C. parapsilosis</i>	
2 (3/day +10)	<i>C. parapsilosis</i>	<i>C. parapsilosis</i>	
2 (4/day +16)	<i>C. parapsilosis</i>	Negative	
2 (5/day +22)	<i>C. parapsilosis</i>	Negative	

3	<i>C. parapsilosis</i>	<i>C. parapsilosis</i>	<i>C. parapsilosis</i> (Urine & BAS)
4	<i>C. parapsilosis</i>	Negative	
5	<i>C. albicans/tropicalis-C. parapsilosis</i>	<i>C. albicans</i>	<i>C. albicans</i> (BAS)
6	<i>C. parapsilosis</i>	<i>C. parapsilosis</i>	<i>C. parapsilosis</i> (Urine)

BAS:Bronchial aspirate

Conclusions:

1. T2Candida® detects five candidemias that were not detected by BC.
2. T2Candida positive result anticipates the BC in 2-5 days.
3. The mean time to negative results was 3.5-4.5 hours for T2Candida compared with 120 h for BC.